

FUJI DRI-CHEM ELECTROLYTE CONTROL QE

■ Intended use

FUJI DRI-CHEM ELECTROLYTE CONTROL QE is a control solution specially designed for FUJI DRI-CHEM system. This product is used to confirm performance using FUJI DRI-CHEM Na-K-Cl slides.

For *in vitro* diagnostic use only.

■ Reagents

This product is an aqueous solution which contains electrolytes, a polymer and other ingredients.

[Reactive ingredients]

Sodium chloride 100 mmol/L

Sodium bicarbonate 30 mmol/L

Potassium biphosphate 4 mmol/L

[Other ingredients]

Polyvinylpyrrolidone

Glycerin

Preservative

■ Test object

Na, K, Cl

■ System applied

1. Analyzer : FUJI DRI-CHEM ANALYZER

2. Slide : FUJI DRI-CHEM SLIDE Na-K-Cl

■ Procedure

1. Wear proper gloves, glasses and other protective gear for your safety.
2. This product should be assayed in the same way as patient specimens.
3. When opening the ampule, take care not to cut a finger on the edge of the glass.
4. Use immediately after opening to avoid concentration.

■ Warnings and precautions

1. Do not use the residual solution in the ampule as it may have concentrated. Discard immediately.
2. This product cannot be applied to other devices not specified by the manufacturer.

■ Storage and shelf life

1. Storage condition : Between 2–8 °C (35.6–46.4 °F) (Do not freeze)
2. Expiry date is printed on the carton.

■ Contents

FUJI DRI-CHEM ELECTROLYTE CONTROL QE : 1 mL x10 ampules



<http://www.fujifilm.com/products/medical/>



FUJIFILM Europe GmbH

Heesenstr. 31, D-40549 Düsseldorf, GERMANY



FUJIFILM Corporation

26-30, Nishazabu 2-Chome, Minato-ku, Tokyo, 106-8620, JAPAN